

MiRo House Becklands Close Boroughbridge North Yorkshire YO51 9NR UNITED KINGDOM

tel: +44(0)1423 320 007





Leeds Test Objects' PIX-13 phantom is designed to be used to check the image quality performance of digital radiography x-ray systems. These tests should be carried out at acceptance to establish a baseline, and on a routine basis, keeping an on-going record of the test results, which will indicate any deterioration in performance.

PIX-13 has been designed in accordance with the **DIN 6868-13** standard and features the following test details:

- Tube axis direction markings
- Limiting spatial resolution (0.6 5.0 LP/mm)
- Low contrast resolution details (6 x 15mm diameter details)
- Dynamic range (7 different thickness Cu steps)
- Unstructured are for placement of dosemeter
- Unstructured area for
- homogeneity and signal adjustment (centre-marked)
- Cassette size alignment markers (radiopaque)
- X-ray to light field alignment markers (radiopaque)

PIX-13 should be used with either PMMA/Cu or Al attenuator plates (available separately).

The phantom is supplied in a soft protective carry bag.

Option to add beam centring tower to the centre of PIX-13

- please contact us for further details and pricing





MiRo House Becklands Close Boroughbridge North Yorkshire YO51 9NR UNITED KINGDOM

tel: +44(0)1423 320 007



Product Detail




fig. 1 PIX-13 phantom

www.leedstestobjects.com




MiRo House Becklands Close Boroughbridge North Yorkshire YO51 9NR UNITED KINGDOM

tel: +44(0)1423 320 007



Product Detail

fig. 2 PIX-13 radiograph



Accessories



An Aluminium attenuator for use with PIX-13 is available separately

www.leeds test objects.com